Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

This is an amended version of report# 26-000838/D008.R000.
Reason: Updated client information.

Customer: Secret Nature

Huntington Beach, CA 92649
United States of America (USA)

Product identity: Live D9 Gummy, Daytrip Strawberry Rhubarb 20mg
Material: Cannabinoid Edible
Laboratory ID: 26-000838-0001
Evidence of Cooling: No
Temp: 12.1 °C
Lot #: 3639NC_011626
Serving Size #1: 8 g

Sample Results

Potency **Method:** J AOAC 2015 V98-6 (mod)[þ] **Batch:** 2600626 **Analyzed:** 01/26/26

Analyte	Result	Units	LOQ	Notes	Serving Size #1 Result	Units	LOQ
CBC	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBC-A	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBC-Total	< LOQ	%	0.0143		< LOQ	mg/8g	1.15
CBD[⊥]	0.00944	%	0.0076		0.755	mg/8g	0.610
CBD-A[⊥]	0.0136	%	0.0076		1.09	mg/8g	0.610
CBD-Total[⊥]	0.0214	%	0.0143		1.71	mg/8g	1.15
CBDV	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBDV-A	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBDV-Total	< LOQ	%	0.0142		< LOQ	mg/8g	1.14
CBE	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBG	0.134	%	0.0076		10.7	mg/8g	0.610
CBG-A	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBG-Total	0.134	%	0.0142		10.7	mg/8g	1.14
CBL	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBL-A	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBL-Total	< LOQ	%	0.0143		< LOQ	mg/8g	1.15
CBN	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
CBT	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ10-THC-9R	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ10-THC-9S	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ10-THC-Total	< LOQ	%	0.0153		< LOQ	mg/8g	1.22
Δ8-THC[⊥]	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ8-THCV	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ9-THC[⊥]	0.266	%	0.0076		21.3	mg/8g	0.610
Δ9-THC-A[⊥]	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ9-THC-Total[⊥]	0.266	%	0.0143		21.3	mg/8g	1.15
Δ9-THCP	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ9-THCV	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ9-THCV-A	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Δ9-THCV-Total	< LOQ	%	0.0142		< LOQ	mg/8g	1.14

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Potency

Method: J AOAC 2015 V98-6 (mod)[þ] **Batch:** 2600626 **Analyzed:** 01/26/26

Analyte	Result	Units	LOQ	Notes	Serving Size #1 Result	Units	LOQ
exo-THC	< LOQ	%	0.0076		< LOQ	mg/8g	0.610
Total Cannabinoids	0.423	%			33.8	mg/8g	

Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed	Method	Status	Notes
Salmonella spp. ⊥	Negative		/25g		2600524	01/24/26	AOAC 2020.02[þ]		
EHEC including STEC ⊥	Negative		/25g		2600525	01/24/26	AOAC 2020.06[þ]		

Solvents

Method: Residual Solvents by HS-GC-MS[þ] **Units** µg/g **Batch** 2600584 **Analyze:** 01/24/26

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane ⊥	< LOQ	380	100	pass		2-Butanol ⊥	< LOQ	5000	200	pass	
2-Ethoxyethanol ⊥	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane) ⊥	< LOQ		200		
2-Methylpentane ⊥	< LOQ		30.0			2-Propanol (IPA) ⊥	< LOQ	5000	200	pass	
2,2-Dimethylbutane ⊥	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane) ⊥	< LOQ		200		
2,3-Dimethylbutane ⊥	< LOQ		30.0			3-Methylpentane ⊥	< LOQ		30.0		
Acetone ⊥	< LOQ	5000	200	pass		Acetonitrile ⊥	< LOQ	410	100	pass	
Benzene ⊥	< LOQ	2.00	1.00	pass		Butanes (sum) ⊥	< LOQ	5000	400	pass	
Cyclohexane ⊥	< LOQ	3880	200	pass		Ethyl acetate ⊥	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether ⊥	< LOQ	5000	200	pass	
Ethylene glycol ⊥	< LOQ	620	200	pass		Ethylene oxide ⊥	< LOQ	50.0	20.0	pass	
Hexanes (sum) ⊥	< LOQ	290	150	pass		Isopropyl acetate ⊥	< LOQ	5000	200	pass	
Isopropylbenzene (Cumene) ⊥	< LOQ	70.0	30.0	pass		m,p-Xylene ⊥	< LOQ		200		
Methanol ⊥	< LOQ	3000	200	pass		Methylene chloride ⊥	< LOQ	600	60.0	pass	
Methylpropane (Isobutane) ⊥	< LOQ		200			n-Butane ⊥	< LOQ		200		
n-Heptane ⊥	< LOQ	5000	200	pass		n-Hexane ⊥	< LOQ		30.0		
n-Pentane ⊥	< LOQ		200			o-Xylene ⊥	< LOQ		200		
Pentanes (sum) ⊥	< LOQ	5000	600	pass		Propane ⊥	< LOQ	5000	200	pass	
Tetrahydrofuran ⊥	< LOQ	720	100	pass		Toluene ⊥	< LOQ	890	100	pass	
Total Xylenes ⊥	< LOQ		400			Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass	

Pesticides

Method: AOAC 2007.01 & EN 15662 (mod) **Units** mg/kg **Batch** 2600666 **Analyze:** 01/28/26

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin ⊥	< LOQ	0.50	0.250	pass		Acephate ⊥	< LOQ	0.40	0.200	pass	
Acequinocyl ⊥	< LOQ	2.0	1.00	pass		Acetamiprid ⊥	< LOQ	0.20	0.100	pass	
Aldicarb ⊥	< LOQ	0.40	0.200	pass		Azoxystrobin ⊥	< LOQ	0.20	0.100	pass	
Bifenazate ⊥	< LOQ	0.20	0.100	pass		Bifenthrin ⊥	< LOQ	0.20	0.100	pass	
Boscalid ⊥	< LOQ	0.40	0.200	pass		Carbaryl ⊥	< LOQ	0.20	0.100	pass	
Carbofuran ⊥	< LOQ	0.20	0.100	pass		Chlorantraniliprole ⊥	< LOQ	0.20	0.100	pass	
Chlorfenapyr ⊥	< LOQ	1.0	0.500	pass		Chlorpyrifos-ethyl ⊥	< LOQ	0.20	0.100	pass	
Clofentezine ⊥	< LOQ	0.20	0.100	pass		Cyfluthrin (sum) ⊥	< LOQ	1.0	0.500	pass	
Cypermethrin (sum) ⊥	< LOQ	1.0	0.500	pass		Daminozide ⊥	< LOQ	1.0	0.500	pass	
Diazinon ⊥	< LOQ	0.20	0.100	pass		Dichlorvos ⊥	< LOQ	1.0	0.500	pass	
Dimethoate ⊥	< LOQ	0.20	0.100	pass		Ethoprophos ⊥	< LOQ	0.20	0.100	pass	
Etofenprox ⊥	< LOQ	0.40	0.200	pass		Etoxazole ⊥	< LOQ	0.20	0.100	pass	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Pesticides

Method: AOAC 2007.01 & EN 15662 (mod) **Units** mg/kg **Batch** 2600666 **Analyze:** 01/28/26

Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Fenoxycarb ⊥	< LOQ	0.20	0.100	pass		Fenpyroximate ⊥	< LOQ	0.40	0.200	pass	
Fipronil ⊥	< LOQ	0.40	0.200	pass		Flonicamid ⊥	< LOQ	1.0	0.400	pass	
Fludioxonil ⊥	< LOQ	0.40	0.200	pass		Hexythiazox ⊥	< LOQ	1.0	0.400	pass	
Imazalil ⊥	< LOQ	0.20	0.100	pass		Imidacloprid ⊥	< LOQ	0.40	0.200	pass	
Kresoxim-methyl ⊥	< LOQ	0.40	0.200	pass		Malathion ⊥	< LOQ	0.20	0.100	pass	
Metalaxyl ⊥	< LOQ	0.20	0.100	pass		Methiocarb ⊥	< LOQ	0.20	0.100	pass	
Methomyl ⊥	< LOQ	0.40	0.200	pass		MGK-264 ⊥	< LOQ	0.20	0.100	pass	
Myclobutanil ⊥	< LOQ	0.20	0.100	pass		Naled ⊥	< LOQ	0.50	0.250	pass	
Oxamyl ⊥	< LOQ	1.0	0.500	pass		Paclobutrazole ⊥	< LOQ	0.40	0.200	pass	
Parathion-methyl ⊥	< LOQ	0.20	0.100	pass		Permethrin ⊥	< LOQ	0.20	0.100	pass	
Phosmet ⊥	< LOQ	0.20	0.100	pass		Piperonyl butoxide ⊥	< LOQ	2.0	1.00	pass	
Prallethrin ⊥	< LOQ	0.20	0.100	pass		Propiconazole ⊥	< LOQ	0.40	0.200	pass	
Propoxur ⊥	< LOQ	0.20	0.100	pass		Pyrethrin I (total) ⊥	< LOQ	1.0	0.500	pass	
Pyridaben ⊥	< LOQ	0.20	0.100	pass		Spinosad ⊥	< LOQ	0.20	0.100	pass	
Spiromesifen ⊥	< LOQ	0.20	0.100	pass		Spirotetramat ⊥	< LOQ	0.20	0.100	pass	
Spiroxamine ⊥	< LOQ	0.40	0.200	pass		Tebuconazole ⊥	< LOQ	0.40	0.200	pass	
Thiacloprid ⊥	< LOQ	0.20	0.100	pass		Thiamethoxam ⊥	< LOQ	0.20	0.100	pass	
Trifloxystrobin ⊥	< LOQ	0.20	0.100	pass							

Metals

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed	Method	Status	Notes
Arsenic ⊥	< LOQ	0.200	mg/kg	0.0184	2600556	01/23/26	AOAC 2013.06 (mod.)þ	pass	
Cadmium ⊥	< LOQ	0.200	mg/kg	0.0184	2600556	01/23/26	AOAC 2013.06 (mod.)þ	pass	
Lead ⊥	< LOQ	0.500	mg/kg	0.0184	2600556	01/23/26	AOAC 2013.06 (mod.)þ	pass	
Mercury ⊥	< LOQ	0.100	mg/kg	0.00921	2600556	01/23/26	AOAC 2013.06 (mod.)þ	pass	

Mycotoxins

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed	Method	Status	Notes
Aflatoxin B1 ⊥	< LOQ		µg/kg	5.00	2600663	01/28/26	Mycotoxins by AOAC 2007.01		
Aflatoxin B2 ⊥	< LOQ		µg/kg	5.00	2600663	01/28/26	Mycotoxins by AOAC 2007.01		
Aflatoxin G1 ⊥	< LOQ		µg/kg	5.00	2600663	01/28/26	Mycotoxins by AOAC 2007.01		
Aflatoxin G2 ⊥	< LOQ		µg/kg	5.00	2600663	01/28/26	Mycotoxins by AOAC 2007.01		
Ochratoxin A ⊥	< LOQ	20.0	µg/kg	5.00	2600663	01/28/26	Mycotoxins by AOAC 2007.01	pass	
Total Aflatoxins	< LOQ	20.0	µg/kg	20.0		01/28/26	Mycotoxins by AOAC 2007.01þ	pass	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number:	26-000838/D008.R001
Report Date:	02/20/2026
ORELAP#:	OR100028
Received:	01/21/26 09:55

Abbreviations
Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Threshold Note: OAR 333-007-0400

þ = ISO/IEC 17025:2017 accredited method.
⊥ = TNI accredited analyte.

Units of Measure

% wt = µg/g divided by 10,000
/25g = Per 25g
µg/g = Microgram per gram
µg/kg = Micrograms per kilogram = parts per billion (ppb)
mg/kg = Milligram per kilogram = parts per million (ppm)
% = Percentage of sample
mg/8g = Milligram per 8g

Approved Signatory

Derrick Tanner
General Manager

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 2 Document ID: 7087
Legacy ID: CFL-E33Effective:

Laboratory Quality Control Results

Residual Solvents **Batch ID:** 2600584

Method Blank					Laboratory Control Sample					
Analyte	**Result**		**LOQ**	**Notes**	**Result**	**Spike**	**Units**	**% Rec**	**Limits**	**Notes**
1,1,1-Trichloroethane	ND	<	5		3.69	5	µg/g	73.8	50-150	
1,2-Dichloroethene, trans-	ND	<	1		0.668	1	µg/g	66.8	50-150	
1,2-Dichloroethene,cis-	ND	<	1		0.704	1	µg/g	70.4	50-150	
1,4-Dioxane	ND	<	100		418	509	µg/g	82.1	60-120	
1-Butanol	ND	<	500		1460	1610	µg/g	90.7	50-150	
2,2-Dimethylbutane	ND	<	30		147	188	µg/g	78.2	60-120	
2,2-Dimethylpropane	ND	<	200		852	956	µg/g	89.1	60-120	
2,3-Dimethylbutane	ND	<	30		147	188	µg/g	78.2	60-120	
2-Butanol	ND	<	200		1310	1640	µg/g	79.9	60-120	
2-Ethoxyethanol	ND	<	30		163	188	µg/g	86.7	60-120	
2-Methylbutane	ND	<	200		1330	1660	µg/g	80.1	60-120	
2-Methylpentane	ND	<	30		153	189	µg/g	81.0	60-120	
2-Propanol	ND	<	200		1400	1680	µg/g	83.3	60-120	
3-Methylpentane	ND	<	30		155	188	µg/g	82.4	60-120	
Acetone	ND	<	200		1400	1670	µg/g	83.8	60-120	
Acetonitrile	ND	<	100		422	511	µg/g	82.6	60-120	
Benzene	ND	<	1		0.696	1	µg/g	69.6	50-150	
Butane	ND	<	200		695	769	µg/g	90.4	60-120	
Chloroform	ND	<	1		0.752	1	µg/g	75.2	50-150	
Cumene	ND	<	30		164	192	µg/g	85.4	60-120	
Cyclohexane	ND	<	200		1390	1650	µg/g	84.2	60-120	
Dichloromethane	ND	<	1		0.671	1	µg/g	67.1	50-150	
DMSO	ND	<	500		1320	1620	µg/g	81.5	50-150	
Ethanol	ND	<	200		1370	1650	µg/g	83.0	60-120	
Ethyl acetate	ND	<	200		1350	1630	µg/g	82.8	60-120	
Ethyl Ether	ND	<	200		1330	1630	µg/g	81.6	60-120	
Ethylbenzene	ND	<	200		845	996	µg/g	84.8	60-120	
Ethylene Glycol	ND	<	200		426	520	µg/g	81.9	60-120	
Ethylene Oxide	ND	<	1		0.688	1	µg/g	68.8	50-150	
Heptane	ND	<	200		1360	1630	µg/g	83.4	60-120	
Hexane	ND	<	30		150	191	µg/g	78.5	60-120	
Isobutane	ND	<	200		682	770	µg/g	88.6	60-120	
Isobutyl Acetate	ND	<	500		1530	1640	µg/g	93.3	50-150	
Isopropyl Acetate	ND	<	200		1430	1660	µg/g	86.1	60-120	
m,p-Xylene	ND	<	200		892	1030	µg/g	86.6	60-120	
Methanol	ND	<	200		1360	1660	µg/g	81.9	60-120	
Methylethylketone	ND	<	500		1510	1620	µg/g	93.2	50-150	
Methylisobutylketone	ND	<	500		1540	1620	µg/g	95.1	50-150	
MTBE	ND	<	500		1430	1610	µg/g	88.8	50-150	
N,N-dimethylformamide	ND	<	150		484	492	µg/g	98.4	50-150	
o-Xylene	ND	<	200		867	996	µg/g	87.0	60-120	
Pentane	ND	<	200		1300	1630	µg/g	79.8	60-120	
Propane	ND	<	200		538	585	µg/g	92.0	60-120	
Tetrahydrofuran	ND	<	100		448	519	µg/g	86.3	60-120	
Toluene	ND	<	100		452	518	µg/g	87.3	60-120	
Triethylamine	ND	<	500		1380	1610	µg/g	85.7	50-150	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 2 Document ID: 7087
Legacy ID: CFL-E33Effective:

QC - Sample Duplicate **Sample ID: 26-000872-0001**

Analyte	SR Result	SD Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
1,1,1-Trichloroethane	ND	ND	5	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethene, trans-	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,2-Dichloroethene,cis-	ND	ND	1	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
1-Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,3-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	846	1120	200	µg/g	27.9	< 20	FAIL	R, E2
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methylethylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

µg/g- Microgram per gram or ppm

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 4 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 **Batch ID: 2600626**

Laboratory Control Sample

Analyte	LCS	Result	Spike	Units	% Rec	Limits	Evaluation	Notes
CBDVA	2	0.0303	0.0295	%	103	80.0 - 120	Acceptable	
CBDV	2	0.0318	0.0311	%	102	80.0 - 120	Acceptable	
CBE	2	0.0344	0.0341	%	101	80.0 - 120	Acceptable	
CBDA	1	0.0293	0.0288	%	102	90.0 - 110	Acceptable	
CBGA	1	0.0304	0.0300	%	101	80.0 - 120	Acceptable	
CBG	1	0.0283	0.0283	%	99.8	80.0 - 120	Acceptable	
CBD	1	0.0273	0.0264	%	103	90.0 - 110	Acceptable	
THCV	2	0.0315	0.0313	%	101	80.0 - 120	Acceptable	
d8THCV	2	0.0322	0.0328	%	98.1	80.0 - 120	Acceptable	
THCVA	2	0.0576	0.0558	%	103	80.0 - 120	Acceptable	
CBN	1	0.0282	0.0277	%	102	80.0 - 120	Acceptable	
exo-THC	2	0.0304	0.0298	%	102	80.0 - 120	Acceptable	
d9THC	1	0.0277	0.0274	%	101	90.0 - 110	Acceptable	
d8THC	1	0.0293	0.0293	%	100	90.0 - 110	Acceptable	
9S-d10THC	1	0.0310	0.0308	%	101	80.0 - 120	Acceptable	
CBL	2	0.0298	0.0292	%	102	80.0 - 120	Acceptable	
9R-d10THC	1	0.0332	0.0330	%	101	80.0 - 120	Acceptable	
CBC	2	0.0319	0.0314	%	102	80.0 - 120	Acceptable	
THCA	1	0.0287	0.0289	%	99.5	90.0 - 110	Acceptable	
CBCA	2	0.0308	0.0306	%	101	80.0 - 120	Acceptable	
CBLA	2	0.0309	0.0306	%	101	80.0 - 120	Acceptable	
d9THCP	2	0.0298	0.0293	%	101	80.0 - 120	Acceptable	
CBT	2	0.0313	0.0316	%	99.0	80.0 - 120	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBDV	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBE	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBDA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBGA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBG	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBD	<LOQ	0.00776	%	< 0.00776	Acceptable	
THCV	<LOQ	0.00776	%	< 0.00776	Acceptable	
d8THCV	<LOQ	0.00776	%	< 0.00776	Acceptable	
THCVA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBN	<LOQ	0.00776	%	< 0.00776	Acceptable	
exo-THC	<LOQ	0.00776	%	< 0.00776	Acceptable	
d9THC	<LOQ	0.00776	%	< 0.00776	Acceptable	
d8THC	<LOQ	0.00776	%	< 0.00776	Acceptable	
9S-d10THC	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBL	<LOQ	0.00776	%	< 0.00776	Acceptable	
9R-d10THC	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBC	<LOQ	0.00776	%	< 0.00776	Acceptable	
THCA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBCA	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBLA	<LOQ	0.00776	%	< 0.00776	Acceptable	
d9THCP	<LOQ	0.00776	%	< 0.00776	Acceptable	
CBT	<LOQ	0.00776	%	< 0.00776	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 4 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6 **Batch ID: 2600626**

Sample Duplicate Sample ID: **26-000835-0001**

Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBDV	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBDA	0.0101	0.0102	0.00770	%	0.416	< 10	Acceptable	
CBGA	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBG	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.00770	%	NA	< 10	Acceptable	
THCV	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
d9THC	0.220	0.217	0.00770	%	1.15	< 10	Acceptable	
d8THC	<LOQ	<LOQ	0.00770	%	NA	< 10	Acceptable	
9S-d10THC	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
9R-d10THC	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBC	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.00770	%	NA	< 10	Acceptable	
CBCA	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
d9THCP	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.00770	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662 **Units:** mg/Kg **Batch ID:** 2600666

Method Blank				Laboratory Control Sample					
Analyte	**Blank Result**	**Blank Limits**	**Notes**	**LCS Result**	**LCS Spike**	**LCS % Rec**	**Limits**		**Notes**
Abamectin	0.000	< 0.070		0.290	0.280	103.5	50.0	150	
Acephate	0.000	< 0.020		0.074	0.080	92.7	60.0	120	
Acequinocyl	0.000	< 0.020		0.107	0.080	133.7	40.0	160	
Acetamiprid	0.000	< 0.020		0.068	0.080	84.4	60.0	120	
Aldicarb	0.000	< 0.100		0.379	0.400	94.7	60.0	120	
Azoxystrobin	0.001	< 0.010		0.037	0.040	92.0	60.0	120	
Bifenazate	0.001	< 0.010		0.041	0.040	103.1	60.0	120	
Bifenthrin	0.000	< 0.100		0.345	0.400	86.3	50.0	150	
Boscalid	0.001	< 0.010		0.037	0.040	92.9	60.0	120	
Carbaryl	0.000	< 0.020		0.069	0.080	86.7	60.0	120	
Carbofuran	0.000	< 0.010		0.038	0.040	94.2	60.0	120	
Chlorantraniliprole	0.000	< 0.010		0.036	0.040	90.4	60.0	120	
Chlorfenapyr	0.000	< 0.050		0.230	0.200	115.1	60.0	120	
Chlorpyrifos	0.000	< 0.010		0.033	0.040	82.4	60.0	120	
Clofentezine	0.001	< 0.010		0.036	0.040	89.5	60.0	120	
Cyfluthrin	0.000	< 0.200		0.643	0.800	80.4	50.0	150	
Cypermethrin	0.009	< 0.200		0.701	0.800	87.6	50.0	150	
Daminozide	0.050	< 0.050		0.186	0.200	93.1	60.0	120	
Diazinon	0.002	< 0.010		0.036	0.040	90.9	60.0	120	
Dichlorvos	0.000	< 0.050		0.179	0.200	89.5	60.0	120	
Dimethoate	0.001	< 0.010		0.036	0.040	89.1	60.0	120	
Ethoprophos	0.001	< 0.010		0.036	0.040	90.4	60.0	120	
Etofenprox	0.002	< 0.010		0.036	0.040	89.9	50.0	150	
Etoxazole	0.000	< 0.010		0.035	0.040	88.7	60.0	120	
Fenoxycarb	0.000	< 0.010		0.041	0.040	103.0	60.0	120	
Fenpyroximate	0.003	< 0.020		0.073	0.080	91.1	60.0	120	
Fipronil	0.000	< 0.010		0.040	0.040	101.2	60.0	120	
Flonicamid	0.000	< 0.020		0.076	0.080	95.1	60.0	120	
Fludioxonil	0.000	< 0.010		0.036	0.040	89.3	50.0	150	
Hexythiazox	0.000	< 0.010		0.036	0.040	88.9	60.0	120	
Imazalil	0.001	< 0.010		0.038	0.040	96.2	60.0	120	
Imidacloprid	0.000	< 0.010		0.035	0.040	88.1	60.0	120	
Kresoxim-methyl	0.000	< 0.010		0.036	0.040	91.0	60.0	120	
Malathion	0.000	< 0.010		0.037	0.040	92.7	60.0	120	
Metalaxyl	0.001	< 0.010		0.035	0.040	88.6	60.0	120	
Methiocarb	0.000	< 0.010		0.036	0.040	90.0	60.0	120	
Methomyl	0.000	< 0.020		0.069	0.080	86.9	60.0	120	
MGK-264	0.002	< 0.050		0.179	0.200	89.6	50.0	150	
Myclobutanil	0.000	< 0.010		0.036	0.040	91.0	60.0	120	
Naled	0.012	< 0.100		0.379	0.400	94.6	50.0	150	
Oxamyl	0.000	< 0.200		0.737	0.800	92.2	60.0	120	
Paclobutrazole	0.001	< 0.010		0.037	0.040	92.2	60.0	120	
Parathion-Methyl	0.010	< 0.030		0.113	0.120	93.8	50.0	150	
Permethrin	0.000	< 0.040		0.132	0.160	82.3	50.0	150	
Phosmet	0.000	< 0.010		0.047	0.040	118.1	50.0	150	
Piperonyl butoxide	0.000	< 0.200		0.754	0.800	94.3	60.0	120	
Prallethrin	0.002	< 0.050		0.173	0.200	86.4	60.0	120	
Propiconazole	0.001	< 0.010		0.036	0.040	89.9	60.0	120	
Propoxur	0.000	< 0.010		0.037	0.040	92.3	60.0	120	
Pyrethrin (Summe)	0.005	< 0.100		0.099	0.100	99.0	60.0	120	
Pyridaben	0.002	< 0.020		0.071	0.080	88.5	50.0	150	
Spinosad	0.000	< 0.100		0.035	0.040	87.7	50.0	150	
Spiromesifen	0.002	< 0.030		0.106	0.120	88.1	60.0	120	
Spirotetramat	0.000	< 0.010		0.039	0.040	96.6	60.0	120	
Spiroxamine	0.000	< 0.010		0.036	0.040	90.3	60.0	120	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.

Columbia LABORATORIES
A Tentamus Company

12423 NE Whitaker Way
Portland, OR 97230
503-254-1794

Report Number: 26-000838/D008.R001
Report Date: 02/20/2026
ORELAP#: OR100028
Received: 01/21/26 09:55

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662 **Units:** mg/Kg **Batch ID:** 2600666

Matrix Spike/Matrix Spike Duplicate Recoveries **Sample ID: 26-000838-0001**

Analyte	Result	MS Res	MSD Res	Spike	RPD%	Limit	MS % Rec	MSD % Rec	Limits	Notes
Abamectin	0.000	0.243	0.225	0.280	7.6%	< 30	86.7%	80.3%	50 - 150	
Acephate	0.000	0.082	0.076	0.080	7.9%	< 30	102.2%	94.4%	50 - 150	
Acequinocyl	0.027	0.103	0.105	0.080	2.4%	< 30	94.8%	97.1%	50 - 150	
Acetamiprid	0.000	0.075	0.072	0.080	3.4%	< 30	93.6%	90.5%	50 - 150	
Aldicarb	0.000	0.355	0.343	0.400	3.3%	< 30	88.7%	85.8%	50 - 150	
Azoxystrobin	0.001	0.033	0.033	0.040	1.6%	< 30	81.5%	80.2%	50 - 150	
Bifenazate	0.001	0.039	0.037	0.040	4.8%	< 30	95.0%	90.5%	50 - 150	
Bifenthrin	0.000	0.345	0.334	0.400	3.4%	< 30	86.3%	83.5%	50 - 150	
Boscalid	0.001	0.034	0.040	0.040	16.9%	< 30	83.2%	98.6%	50 - 150	
Carbaryl	0.000	0.071	0.063	0.080	12.4%	< 30	88.7%	78.3%	50 - 150	
Carbofuran	0.000	0.033	0.034	0.040	2.0%	< 30	82.6%	84.3%	50 - 150	
Chlorantraniliprole	0.000	0.032	0.034	0.040	5.7%	< 30	79.6%	84.2%	50 - 150	
Chlorfenapyr	0.000	0.345	0.428	0.200	21.4%	< 30	172.7%	214.2%	50 - 150	Q
Chlorpyrifos	0.000	0.035	0.032	0.040	10.1%	< 30	88.5%	80.0%	50 - 150	
Clofentezine	0.001	0.035	0.034	0.040	4.4%	< 30	87.0%	83.3%	50 - 150	
Cyfluthrin	0.000	0.666	0.650	0.800	2.4%	< 30	83.3%	81.3%	30 - 150	
Cypermethrin	0.009	0.760	0.761	0.800	0.1%	< 30	93.8%	94.0%	50 - 150	
Daminozide	0.112	0.240	0.240	0.200	0.4%	< 30	63.7%	64.0%	30 - 150	
Diazinon	0.002	0.035	0.036	0.040	1.6%	< 30	83.7%	85.1%	50 - 150	
Dichlorvos	0.000	0.170	0.174	0.200	2.1%	< 30	85.2%	87.0%	50 - 150	
Dimethoate	0.001	0.036	0.035	0.040	1.8%	< 30	86.9%	85.3%	50 - 150	
Ethoprophos	0.001	0.035	0.036	0.040	2.1%	< 30	85.7%	87.5%	50 - 150	
Etofenprox	0.000	0.035	0.037	0.040	5.8%	< 30	86.4%	91.6%	50 - 150	
Etoxazole	0.000	0.035	0.034	0.040	4.7%	< 30	88.2%	84.1%	50 - 150	
Fenoxycarb	0.001	0.036	0.032	0.040	12.2%	< 30	87.2%	77.2%	50 - 150	
Fenpyroximate	0.003	0.071	0.071	0.080	1.0%	< 30	85.3%	86.2%	50 - 150	
Fipronil	0.000	0.033	0.038	0.040	14.1%	< 30	82.4%	94.9%	50 - 150	
Flonicamid	0.000	0.074	0.067	0.080	10.0%	< 30	93.0%	84.2%	50 - 150	
Fludioxonil	0.000	0.029	0.033	0.040	12.3%	< 30	72.3%	81.8%	50 - 150	
Hexythiazox	0.000	0.041	0.039	0.040	4.7%	< 30	102.1%	97.4%	50 - 150	
Imazalil	0.001	0.034	0.038	0.040	12.0%	< 30	80.7%	91.0%	50 - 150	
Imidacloprid	0.000	0.036	0.033	0.040	8.6%	< 30	89.6%	82.2%	50 - 150	
Kresoxim-methyl	0.000	0.032	0.034	0.040	6.0%	< 30	79.0%	83.8%	50 - 150	
Malathion	0.000	0.035	0.034	0.040	2.8%	< 30	87.1%	84.7%	50 - 150	
Metalaxyl	0.000	0.034	0.033	0.040	4.7%	< 30	85.8%	81.9%	50 - 150	
Methiocarb	0.000	0.037	0.034	0.040	6.9%	< 30	90.3%	84.3%	50 - 150	
Methomyl	0.000	0.066	0.072	0.080	9.0%	< 30	82.4%	90.1%	50 - 150	
MGK-264	0.000	0.173	0.172	0.200	0.6%	< 30	86.6%	86.1%	50 - 150	
Myclobutanil	0.000	0.034	0.033	0.040	1.4%	< 30	83.7%	82.5%	50 - 150	
Naled	0.000	0.372	0.388	0.400	4.4%	< 30	92.9%	97.0%	50 - 150	
Oxamyl	0.000	0.715	0.693	0.800	3.2%	< 30	89.4%	86.6%	50 - 150	
Paclobutrazole	0.000	0.035	0.034	0.040	2.8%	< 30	88.6%	86.2%	50 - 150	
Parathion-Methyl	0.015	0.110	0.105	0.120	5.0%	< 30	79.3%	75.4%	30 - 150	
Permethrin	0.000	0.136	0.154	0.160	12.8%	< 30	84.8%	96.4%	50 - 150	
Phosmet	0.000	0.051	0.046	0.040	9.9%	< 30	128.3%	116.2%	50 - 150	
Piperonyl butoxide	0.000	0.725	0.741	0.800	2.2%	< 30	90.7%	92.7%	50 - 150	
Prallethrin	0.000	0.168	0.172	0.200	2.3%	< 30	83.9%	85.8%	50 - 150	
Propiconazole	0.000	0.033	0.037	0.040	12.2%	< 30	82.9%	93.7%	50 - 150	
Propoxur	0.000	0.038	0.036	0.040	4.8%	< 30	95.0%	90.6%	50 - 150	
Pyrethrin (Summe)	0.004	0.111	0.099	0.100	12.1%	< 30	107.8%	95.5%	50 - 150	
Pyridaben	0.002	0.078	0.077	0.080	2.0%	< 30	95.3%	93.4%	50 - 150	
Spinosad	0.000	0.034	0.036	0.040	7.9%	< 30	83.8%	90.6%	50 - 150	
Spiromesifen	0.000	0.099	0.106	0.120	7.8%	< 30	82.1%	88.7%	50 - 150	
Spirotetramat	0.000	0.036	0.036	0.040	0.2%	< 30	90.0%	89.8%	50 - 150	
Spiroxamine	0.000	0.034	0.036	0.040	3.3%	< 30	85.9%	88.8%	50 - 150	

Test results relate only to the parameters tested and to the samples as received by the laboratory. Test results meet all requirements of Columbia Laboratories quality assurance plan unless otherwise noted. This report shall not be reproduced, except in full, without the written consent of this laboratory. Samples will be retained for a maximum of 30 days from the receipt date unless prior arrangements have been made.